



Medline Industries, Inc.  
Three Lakes Dr.  
Northfield, IL 60093  
1.630.865.6288  
1.800.MEDLINE (633.5463)

## CERTIFICATE OF ANALYSIS

**MCHEM145 - Isopropyl Alcohol 70% 1 gal  
(MTO)**

Lot Number: 230833

Expiration Date: 2027-June-30

### COMPONENTS

Test	Specifications	Results
Appearance	Clear	Clear
Assay (by volume)	$\geq 68.0$ and $\leq 72.0$	70.9
Water (by volume)	$\geq 28.0$ and $\leq 32.0$	29.1
Specific Gravity @ 20C	$\geq 0.865$ and $\leq 0.879$	0.869
Color (APHA)	$\geq 0$ and $\leq 5$	5
Identification Conforms to IR Spectrum	Pass/Fail	Pass

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control - This report is electronically produced and is valid without signature.

Source: SOL-36700/231006  
Form: WM2103\_V5



**Medline Industries, Inc.**  
Three Lakes Dr.  
Northfield, IL 60093  
1.630.865.6288  
1.800.MEDLINE (633.5463)

## CERTIFICATE OF ANALYSIS

**MCHEMPF20N - Prefill, Medline 10%  
NBF, 20mL,96 Vials/cs**

Lot Number: 230144

Expiration Date: 2028-June-30

### COMPONENTS

Test	Specifications	Results
Color	Clear	Clear
pH @ 21 C - 28 C	$\geq 6.8$ and $\leq 7.2$	6.9
Specific Gravity (g/cm <sup>3</sup> )	$\geq 1.014$ and $\leq 1.019$	1.014
Formalin Assay (%)	$\geq 9.0$ and $\leq 11.0$	9.1

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control - This report is electronically produced and is valid without signature.

Source: 28600/233172  
Form: WM2103\_V5



## CERTIFICATE OF ANALYSIS

NB16240 - Prefill Statlab NBF 500 mL 6  
Vials/Case

Lot Number: 231222

Expiration Date: 2028-June-30

### COMPONENTS

Test	Specifications	Results
Color	Clear	Clear
pH @ 21 C - 28 C	$\geq 6.8$ and $\leq 7.2$	6.9
Specific Gravity (g/cm <sup>3</sup> )	$\geq 1.014$ and $\leq 1.019$	1.014
Formalin Assay (%)	$\geq 9.0$ and $\leq 11.0$	9.7

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control - This report is electronically produced and is valid without signature.

Source: 28600/232751  
Form: WM2103\_V5