



**Medline Industries, Inc.**  
Three Lakes Dr.  
Northfield, IL 60093  
1.630.865.6288  
1.800.MEDLINE (633.5463)

## CERTIFICATE OF ANALYSIS

**MCHEM145 - Isopropyl Alcohol 70% 1 gal  
(MTO)**

Lot Number: 247376

Expiration Date: 2027-December-31

### COMPONENTS

Test	Specifications	Results
Appearance	Clear	Clear
Assay (by volume)	$\geq 68.0$ and $\leq 72.0$	70.2
Water (by volume)	$\geq 28.0$ and $\leq 32.0$	29.8
Specific Gravity @ 20C	$\geq 0.865$ and $\leq 0.879$	0.872
Color (APHA)	$\geq 0$ and $\leq 5$	4
Identification Conforms to IR Spectrum	Pass/Fail	Pass

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control - This report is electronically produced and is valid without signature.

Source: SOL-36700/246806  
Form: WM2103\_V5