

April 9, 2015

MEDLINE INDUSTRIES INC
1 MEDLINE PLACE
MUNDELEIN IL 60060

Re: Assigned HCPCS Codes for DME Billing

Xref: 38449027

WALKER PNEUMATIC DELUXE XS EA	MEDLINE INDUSTRIES INC	ORT28610XS	L4360 OR L4361
WALKER PNEUMATIC DELUXE SM EA	MEDLINE INDUSTRIES INC	ORT28610S	L4360 OR L4361
WALKER PNEUMATIC DELUXE MED EA	MEDLINE INDUSTRIES INC	ORT28610M	L4360 OR L4361
WALKER PNEUMATIC DELUXE LRG EA	MEDLINE INDUSTRIES INC	ORT28610L	L4360 OR L4361
WALKER PNEUMATIC DELUXE XL EA	MEDLINE INDUSTRIES INC	ORT28610XL	L4360 OR L4361
ANKLE WALKER DELUXE PNEUMATIC XS EA	MEDLINE INDUSTRIES INC	ORT28620XS	L4360 OR L4361
ANKLE WALKER DELUXE PNEUMATIC S EA	MEDLINE INDUSTRIES INC	ORT28620S	L4360 OR L4361
ANKLE WALKER DELUXE PNEUMATIC M EA	MEDLINE INDUSTRIES INC	ORT28620M	L4360 OR L4361
ANKLE WALKER DELUXE PNEUMATIC L EA	MEDLINE INDUSTRIES INC	ORT28620L	L4360 OR L4361
ANKLE WALKER DELUXE PNEUMATIC XL EA	MEDLINE INDUSTRIES INC	ORT28620XL	L4360 OR L4361

Dear Kristen Freund:

The Pricing, Data Analysis, and Coding (PDAC) Contractor has reviewed the product(s) listed above and has approved the listed Healthcare Common Procedure Coding System (HCPCS) code(s) for billing the four Durable Medical Equipment Medicare Administrative Contractors (DME MACs).

The PDAC Contractor provides coding assistance to manufacturers to ensure proper coding of Durable Medical Equipment, Prosthetics, Orthotics, and Supplies (DMEPOS). The PDAC publishes coding decisions based on the coding guidelines established by the Local Coverage Determinations (LCDs) and associated Policy Articles and any related Advisory Articles

established by the DME MACs. All products submitted to the PDAC for a coding verification review are examined by coders and professionals following a formal, standardized process.

The PDAC has reviewed the above listed product(s). Based on this review and application of DME MAC policy, the HCPCS code(s) listed below should be used when billing the DME MACs:

L4360 - Walking Boot, Pneumatic And/Or Vacuum, With Or Without Joints, With Or Without Interface Material, Prefabricated Item That Has Been Trimmed, Bent, Molded, Assembled, Or Otherwise Customized To Fit A Specific Patient By An Individual With Expertise

L4361 - Walking Boot, Pneumatic And/Or Vacuum, With Or Without Joints, With Or Without Interface Material, Prefabricated, Off-The-Shelf

This decision applies to the application we received on January 30, 2015. If information submitted in that application has changed or were to change, it could impact our decision. Therefore, a new application would need to be submitted for HCPCS coding verification review. The coding assigned in this decision letter will be available on the Product Classification List (PCL) on the Durable Medical Equipment Coding System (DMECS) within ten (10) working days from the letter's date. The DMECS can be accessed on the PDAC website, www.dmepdac.com. Please take the time to verify that this coding decision is correctly reflected in DMECS.

If you disagree with this decision, you may request a reconsideration within 45 days of the letter's date and provide evidence to substantiate a reconsideration of PDAC's original coding determination. To request a reconsideration, complete the Reconsideration Request form located on the PDAC website at <https://www.dmepdac.com/review/requesting.html>. If your request for a reconsideration is made after the 45-day time frame, it will require a new application and documentation to support the request.

It is the responsibility of manufacturers and distributors to notify the PDAC immediately of any changes involving their products, as listed on the PCL on DMECS. Further information for requesting updates to the PCL can be found on the PDAC website at <https://www.dmepdac.com/review/notifying.html>. It is also the responsibility of manufacturers and distributors to assure their websites and product marketing materials accurately reflect the product reviewed by the PDAC and the coding decision assigned.

An assignment of the HCPCS code(s) to product(s) is not an approval or endorsement of the product(s) by Medicare or Noridian Healthcare Solutions; nor does it imply or guarantee claim reimbursement or coverage.

If you have questions about policy, claim coverage or reimbursement, please contact the DME MAC for your jurisdiction.

For other questions, contact the PDAC Contact Center at the address listed above or by telephone at (877) 735-1326. The Contact Center is open Monday through Friday from 8:30 a.m. to 4 p.m. CT.

Sincerely,

PDAC
Noridian Healthcare Solutions, LLC
www.dmepdac.com