



September 30, 2021

KRISTEN TREASURE  
MEDLINE INDUSTRIES INC  
3 LAKES DR  
NORTHFIELD, IL 60093

**Document Control Number (DCN): 21204003000004**

Manufacturer Name	Product Name	Model Number	Assigned HCPCS Code(s)
MEDLINE INDUSTRIES INC	EXCEL K3 WHEELCHAIR	K3166N34S	K0006
MEDLINE INDUSTRIES INC	EXCEL K3 WHEELCHAIR	K3166N34E	K0006+K0195
MEDLINE INDUSTRIES INC	EXCEL K3 WHEELCHAIR	K3186N34S	K0006
MEDLINE INDUSTRIES INC	EXCEL K3 WHEELCHAIR	K3186N34E	K0006+K0195
MEDLINE INDUSTRIES INC	EXCEL K3 WHEELCHAIR	K3206N34S	K0006+E2201
MEDLINE INDUSTRIES INC	EXCEL K3 WHEELCHAIR	K3206N34E	K0006+K0195+E2201

Dear KRISTEN TREASURE,

The Pricing, Data Analysis, and Coding (PDAC) Contractor has reviewed the product(s) listed above and has approved the listed Healthcare Common Procedure Coding System (HCPCS) code(s) for billing the four Durable Medical Equipment Medicare Administrative Contractors (DME MACs).

The PDAC Contractor provides coding assistance to manufacturers to ensure proper coding of Durable Medical Equipment, Prosthetics, Orthotics, and Supplies (DMEPOS). The PDAC



publishes coding decisions based on the coding guidelines established by the Local Coverage Determinations (LCDs) and associated Policy Articles and any related Advisory Articles established by the DME MACs. All products submitted to the PDAC for a coding verification review are examined by coders and professionals following a formal, standardized process.

Based on this review and application of DME MAC policy, the HCPCS code(s) listed below should be used when billing the DME MACs:

K0006 HEAVY DUTY WHEELCHAIR

K0195 ELEVATING LEG RESTS, PAIR (FOR USE WITH CAPPED RENTAL WHEELCHAIR BASE)

E2201 MANUAL WHEELCHAIR ACCESSORY, NONSTANDARD SEAT FRAME, WIDTH GREATER THAN OR EQUAL TO 20 INCHES AND LESS THAN 24 INCHES

If you disagree with this decision, you may request a reconsideration within 45 days of the letter's date and provide evidence to substantiate a reconsideration of PDAC's original coding determination. To request a reconsideration, complete the Reconsideration Request form located on the PDAC website at [www.dmepdac.com](http://www.dmepdac.com). If your request for a reconsideration is made after the 45-day time frame, it will require a new application and documentation to support the request.

It is the responsibility of manufacturers and distributors to notify the PDAC immediately of any changes involving their products, as listed on the Product Classification List (PCL) on the Durable Medical Equipment Coding System (DMECS). Further information for requesting updates to the PCL can be found on the PDAC website at [www.dmepdac.com](http://www.dmepdac.com). It is also the responsibility of manufacturers and distributors to assure their websites and product marketing materials accurately reflect the product reviewed by the PDAC and the coding decision assigned.

An assignment of the HCPCS code(s) to product(s) is not an approval or endorsement of the product(s) by Medicare or Palmetto GBA; nor does it imply or guarantee claim reimbursement or coverage.

If you have questions, please contact the PDAC HCPCS Helpline at (877) 735-1326 during the hours of 9:30 a.m. to 5:00 p.m. ET, Monday through Friday. You may also visit our [website](#) to chat with one of our representatives or select the Contact Us button at the top of the page for email, FAX or postal mail information.

Sincerely,

Pricing, Data Analysis, and Coding (PDAC)  
Palmetto GBA, LLC  
[www.dmepdac.com](http://www.dmepdac.com)