



Release Date: 5/30/18

REF	Product Name
GTIN	

08P7001 00380740121709	<i>Alinity c ICT Urine Calibrator Kit</i>
---	---

Components:

08P70-01 HIGH	Alinity c ICT Urine Calibrator High
08P70-01 LOW	Alinity c ICT Urine Calibrator Low

Abbott Customers:

For additional information, please contact your Abbott Customer Support Center Representative by calling 1-800-527-1869, 1-800-323-9100, or 1-800-235-5396.

Abbott employees:

For additional information relative to the content of the SDSs, please contact your local Safety Representative.



Abbott

© Abbott Laboratories Release date 05/30/2018

Safety Data Sheet

Last alteration on 05/03/2018

1 Identification

· **Product name:** Alinity c ICT Urine Calibrator High

· **ADD List number:** 08P70-01 HIGH

· **Application of the substance / mixture:** For In Vitro Diagnostic Use

· **Manufacturer / Supplier:**

Abbott Diagnostics
100 Abbott Park Road
Abbott Park, IL 60064-3500

Phone: 1-877-4 ABBOTT

· **Department issuing SDS:** Abbott Diagnostics Environmental Health and Safety

· **Emergency telephone number**

Contact the CHEMTREC® Emergency Call Center for assistance with transportation or hazardous materials emergencies (24 hours/day, 7 days/week). Refer to Abbott customer number 675805.

- Telephone (800) 424-9300 (toll-free) if you are calling from within the United States, Canada, Puerto Rico and the Virgin Islands.

- Telephone +1 (703) 527-3887, the international and maritime number (collect calls accepted), if you are calling from outside the United States or from a ship at sea.

2 Hazard(s) identification

· **Classification of the substance or mixture**

This product has been evaluated per the classification criteria in the Globally Harmonized System of Classification and Labelling of Chemicals (GHS). This product does not meet the criteria for classification in accordance with the GHS.

· **Label elements**

· **GHS label elements:** none

· **Hazard pictograms:** none

· **Signal word:** none

· **Hazard statements:** none

· **Routes of Exposure:**

- Skin: No adverse effects expected when used as directed.

- Eye: No adverse effects expected when used as directed.

- Inhalation: No adverse effects expected when used as directed.

- Ingestion: No adverse effects expected when used as directed.

· **Hazard Overview**

· **Health:** No adverse effects expected if used as directed.

· **Fire:** Noncombustible

· **Reactivity:** Minimal hazard - Stable, even in a fire. Not reactive with water. Not an oxidizer.

3 Composition/information on ingredients

· **Chemical characterization:** Mixture of chemical and/or biological substances for in vitro diagnostic use.



Abbott

© Abbott Laboratories Release date 05/30/2018

Safety Data Sheet

Last alteration on 05/03/2018

Product name: Alinity c ICT Urine Calibrator High

- **Hazardous chemical ingredients per U.S. OSHA criteria (29 CFR 1910.1200 Hazard Communication):** While this product is not considered hazardous by the criteria in 29 CFR 1910.1200, important information regarding the safe handling, transport and disposal of this product is contained in this SDS.

4 First-aid measures

- **After inhalation:** Remove from source of exposure. If irritation or signs of toxicity occur, seek medical attention.
- **After skin contact:**
Take off any clothing that the product touched.
Rinse skin with running water for 15 to 20 minutes. Seek medical attention if irritation or signs of toxicity occur.
- **After eye contact:**
Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. If eye irritation persists, get medical advice/attention. Wash hands after handling.
- **After swallowing:** Rinse mouth with water. If irritation or signs of toxicity occur, seek medical attention.
- **Information for Medical Personnel**
 - **Most important symptoms and effects, both acute and delayed:** None expected
 - **Medical conditions aggravated by exposure:** None known

5 Fire-fighting measures

- **Suitable extinguishing agents**
Dry chemical, carbon dioxide (CO₂), water spray or regular foam.
 - Caution: CO₂ will displace air in confined spaces and may cause an oxygen-deficient atmosphere.
 - For larger fires: There are no unique chemical or reactivity hazards that would impact firefighting decisions related to this product. Use firefighting measures that suit the environment.
- **Special hazards arising from the substance or mixture**
There are no unique chemical or reactivity hazards that would impact firefighting decisions due to the chemicals in this product.
- **Protective equipment**
For large fires, wear appropriate heat- and flame-resistant personal protective equipment and a NFPA/NIOSH approved positive-pressure, self-contained breathing apparatus.

6 Accidental release measures

- **Personal precautions, protective equipment and emergency procedures**
Minimize exposure by using appropriate personal protective equipment as listed in Section 8. Stop leak if possible. Keep unprotected persons away.
- **Environmental precautions**
Prevent liquid and vapor from entering sewage system, storm drains, surface waters, and soil.



Abbott

© Abbott Laboratories Release date 05/30/2018

Safety Data Sheet

Last alteration on 05/03/2018

Product name: Alinity c ICT Urine Calibrator High

· **Methods and material for containment and cleaning up**

Blot up small volumes of spilled or spattered product with paper towels or similar materials.

- Contain larger spills by placing absorbants around the outside edges of the spill. Absorb with any material suitable for water-based liquids - e.g. paper towels, universal sorbents, sand, diatomite, sawdust, etc.

Clean the affected area. Suitable cleaners are:

- warm water and detergent or similar cleansing agent

Dispose of spilled and contaminated material in accordance with Federal, State, and Local regulations. See Section 13 for information that may impact disposal of materials contaminated with this product.

· **Reference to other sections**

See Section 7 for information on safe handling.

See Section 8 for information on personal protection equipment.

See Section 13 for disposal information.

7 Handling and storage

· **Precautions for safe handling:**

Avoid direct contact with material. If handled, wash thoroughly. Practice general safety precautions.

· **Information about protection against explosions and fires:** The product is not flammable.

· **Requirements to be met by storerooms and receptacles:**

Store only in the original container.

Refer to the package insert or product label for additional information on storage conditions for product quality.

· **Information about storage in one common storage facility:** Store in original packaging.

· **Further information about storage conditions:** Protect from heat and direct sunlight.

8 Exposure controls/personal protection

· **Components with Occupational Exposure Limits**

The product does not contain any hazardous ingredients with occupational exposure limits established by OSHA, ACGIH, or NIOSH.

· **General protective and hygienic measures:**

Always maintain good housekeeping and follow general precautionary measures. Do not eat, drink or store food and beverages in areas where chemicals or specimens are used. Wash hands before breaks, after handling reagents and specimens, and at the end of the workshift.

· **Breathing equipment:**



Product name: Alinity c ICT Urine Calibrator High

Normal use and storage of product - respiratory protection is not necessary if room is well ventilated.

Small-volume spills (e.g. small enough to clean up with a paper towel or small sorbent pad) - respiratory protection should not be necessary if room is well ventilated.

Other unusual conditions (e.g. volume spilled too big to clean up with materials in arm's reach) - Use appropriate NIOSH-approved air-purifying respirator if airborne chemical concentrations may exceed the exposure limit (if any) listed above.

Hazardous Materials Emergencies or Firefighting - use NIOSH/NFPA-approved respiratory protection.

· **Hand protection:**

Wear impervious gloves if hand contact with the material is anticipated. Dispose of contaminated gloves after use in accordance with applicable laws and good laboratory practices.

· **Material of gloves and breakthrough time of the glove material:**

The glove material must be suitable for use in a microbiological laboratory and have a measured breakthrough time of at least 30 minutes, such as those with a Class 2 protection index per EN374 (or equivalent standard applicable in your region). NOTE: This recommendation applies only to the product stated in this Safety Data Sheet. When dissolving in or mixing with other substances, contact the supplier of approved gloves.

· **Eye protection:**

Wear safety glasses or other protective eyewear. If splash potential exists, wear full face shield or goggles.

· **Body protection:**

Normal use: protect personal clothing from splatters and small spills. Wear a laboratory coat (or other protective clothing required by your institution).

Larger spills (e.g. that can saturate cloth): wear appropriate water-repellant covering over clothing.

9 Physical and chemical properties

· **General Information**

· **Form:** Solution
· **Color:** Light yellow

· **Odor:** Odorless

· **pH-value at 20 °C (68 °F)** 7
· **Melting point/Melting range:** Not determined
· **Boiling point/Boiling range:** Not determined

· **Flash point** Not applicable
· **Flammability (solid, gaseous)** Not applicable
· **Auto igniting** Product is not self-igniting.

· **Danger of explosion** Product does not present an explosion hazard.

· **Explosion limits**
· **Lower:** Not determined
· **Upper:** Not determined

· **Density at 20 °C (68 °F)** 1 g/cm³ (8.345 lbs/gal)



Product name: Alinity c ICT Urine Calibrator High

· Evaporation rate:	Not determined
· Solubility in / Miscibility with	
· Water:	Fully miscible
· Dynamic:	Not determined
· Water:	98.0 %
· Solids content:	0.0 %

10 Stability and reactivity

- **Thermal decomposition / conditions to be avoided**
No decomposition if used and stored according to specifications.
- **Possibility of hazardous reactions:** No dangerous reactions known.
- **Conditions to avoid:** No further relevant information available.
- **Incompatible materials:** No further relevant information available.
- **Hazardous decomposition products:** No dangerous decomposition products known.

11 Toxicological information

- **Acute toxicity**
 - **LD50/LC50 values for hazardous ingredients per OSHA criteria:**
 - **Ingredients (100% pure substance/s):** Not applicable.
 - **Primary toxicological effects of the final product:**
 - **Skin irritation:** No irritant effect.
 - **Eye irritation:** No irritant effect.
 - **Sensitization:** No sensitizing effects known.
- **Additional toxicological information:** None
 - **Carcinogenic categories**

· IARC (International Agency for Research on Cancer)
None of the ingredients is listed.

· NTP (National Toxicology Program)
None of the ingredients is listed.

· OSHA-Ca (Occupational Safety & Health Administration)
None of the ingredients is listed.

- **Target organs/systems:** Unknown



Product name: Alinity c ICT Urine Calibrator High

12 Ecological information

- **Aquatic toxicity:** No further relevant information available.
- **Additional ecological information**
 - **General notes:** Not known to be a water pollutant.
- **Results of PBT and vPvB assessment**
 - **PBT:** Not applicable
 - **vPvB:** Not applicable

13 Disposal considerations

- **Recommendation for disposal of unused product:**
Dispose in accordance with federal, state and local regulations.
- **Recommendation for disposal of packaging:**
Non-contaminated packaging may be used for recycling. Refer to applicable local regulations and institutional policies.
For disposal of contaminated packaging, refer to applicable local regulations and institutional policies.
- **Recommended cleansing agent:** Water with cleansing agents, if necessary.

14 Transport information

· DOT, ADN, IMDG, IATA none

· UN proper shipping name

· DOT, ADR, ADN, IMDG, IATA none

· Transport hazard class(es)

· DOT, ADR, ADN, IMDG, IATA

· Class none

· DOT, IMDG, IATA none

· Environmental hazards

· Marine pollutant: No

· Additional information

· DOT

· Remarks: Not restricted for transportation.

· ADR

· Remarks: Not restricted for transportation.

· IMDG

· Remarks: Not restricted for transportation.



Abbott

© Abbott Laboratories Release date 05/30/2018

Safety Data Sheet

Last alteration on 05/03/2018

Product name: Alinity c ICT Urine Calibrator High

· IATA

· **Remarks:** Not restricted for transportation.

15 Regulatory information

· **SARA (Superfund Amendments and Reauthorization Act of 1986 - USA):**

· **Section 302/304 (40CFR355.30 / 40CFR355.40):**

The product does not contain listed substances.

· **Section 313 (40CFR372.65):**

The product does not contain listed substances.

· **California Proposition 65 (USA):**

· **Chemicals known to cause cancer:**

The product does not contain listed substances.

· **Chemicals known to cause female reproductive toxicity:**

None of the ingredients is listed.

· **Chemicals known to cause male reproductive toxicity:**

None of the ingredients is listed.

· **Chemicals known to cause developmental reproductive toxicity:**

None of the ingredients is listed.

16 Other information

The information and recommendations contained herein are based upon information or tests believed to be reliable. Abbott Laboratories does not guarantee the accuracy or completeness of this information or recommendations contained herein, NOR SHALL ANY OF THIS INFORMATION CONSTITUTE A WARRANTY, WHETHER EXPRESSED OR IMPLIED, AS TO THE SAFETY OF THE GOODS, THE MERCHANTABILITY OF THE GOODS, OR THE FITNESS OF THE GOODS FOR A PARTICULAR PURPOSE.

This information is not a substitute for the advice of a health care professional, nor is it a recommendation for any particular course of treatment. It is not intended to supplement, modify or supersede any information (e.g. labeling and package inserts) provided with respect to the medical use of the product. Abbott Laboratories assumes no responsibility for results obtained or for incidental or consequential damages, including lost profits, arising from the use of these data. No warranty against infringement of any patent, copyright or trademark is made or implied.

· **Department issuing SDS**

- Abbott Diagnostics Safety, Health and Environmental Assurance
Department 0571

· **Contact**

- General information about this product:
Abbott Diagnostics
Technical Support
100 Abbott Park Road
Abbott Park, IL 60064-3500



Abbott

© Abbott Laboratories Release date 05/30/2018

Safety Data Sheet

Last alteration on 05/03/2018

Product name: Alinity c ICT Urine Calibrator High

Phone: 1-877-4 ABBOTT

· **Date of preparation / last revision** 05/30/2018 / 2

· **Abbreviations and acronyms:**

RID: Règlement international concernant le transport des marchandises dangereuses par chemin de fer (Regulations Concerning the International Transport of Dangerous Goods by Rail)

ICAO: International Civil Aviation Organisation

ADR: Accord européen sur le transport des marchandises dangereuses par Route (European Agreement concerning the International Carriage of Dangerous Goods by Road)

IMDG: International Maritime Code for Dangerous Goods

DOT: US Department of Transportation

IATA: International Air Transport Association

EINECS: European Inventory of Existing Commercial Chemical Substances

ELINCS: European List of Notified Chemical Substances

CAS: Chemical Abstracts Service (Division of the American Chemical Society)

LC50: Lethal concentration, 50 percent

LD50: Lethal dose, 50 percent

PBT: persistent, bioaccumulative and toxic

vPvB: very persistent and very bioaccumulative

OSHA: Occupational Safety & Health

TLV: Threshold Limit Value

PEL: Permissible Exposure Limit

REL: Recommended Exposure Limit

· *** Sections marked with an asterisk (*) have been altered since the previous version.**



Abbott

© Abbott Laboratories Release date 05/30/2018

Safety Data Sheet

Last alteration on 05/03/2018

1 Identification

· **Product name:** Alinity c ICT Urine Calibrator Low

· **ADD List number:** 08P70-01 LOW

· **Application of the substance / mixture:** For In Vitro Diagnostic Use

· **Manufacturer / Supplier:**

Abbott Diagnostics
100 Abbott Park Road
Abbott Park, IL 60064-3500

Phone: 1-877-4 ABBOTT

· **Department issuing SDS:** Abbott Diagnostics Environmental Health and Safety

· **Emergency telephone number**

Contact the CHEMTREC® Emergency Call Center for assistance with transportation or hazardous materials emergencies (24 hours/day, 7 days/week). Refer to Abbott customer number 675805.

- Telephone (800) 424-9300 (toll-free) if you are calling from within the United States, Canada, Puerto Rico and the Virgin Islands.

- Telephone +1 (703) 527-3887, the international and maritime number (collect calls accepted), if you are calling from outside the United States or from a ship at sea.

2 Hazard(s) identification

· **Classification of the substance or mixture**

This product has been evaluated per the classification criteria in the Globally Harmonized System of Classification and Labelling of Chemicals (GHS). This product does not meet the criteria for classification in accordance with the GHS.

· **Label elements**

· **GHS label elements:** none

· **Hazard pictograms:** none

· **Signal word:** none

· **Hazard statements:** none

· **Routes of Exposure:**

- Skin: No adverse effects expected when used as directed.
- Eye: No adverse effects expected when used as directed.
- Inhalation: No adverse effects expected when used as directed.
- Ingestion: No adverse effects expected when used as directed.

· **Hazard Overview**

· **Health:** No adverse effects expected if used as directed.

· **Fire:** Noncombustible

· **Reactivity:** Minimal hazard - Stable, even in a fire. Not reactive with water. Not an oxidizer.

3 Composition/information on ingredients

· **Chemical characterization:** Mixture of chemical and/or biological substances for in vitro diagnostic use.



Abbott

© Abbott Laboratories Release date 05/30/2018

Safety Data Sheet

Last alteration on 05/03/2018

Product name: Alinity c ICT Urine Calibrator Low

- **Hazardous chemical ingredients per U.S. OSHA criteria (29 CFR 1910.1200 Hazard Communication):** While this product is not considered hazardous by the criteria in 29 CFR 1910.1200, important information regarding the safe handling, transport and disposal of this product is contained in this SDS.

4 First-aid measures

- **After inhalation:** Remove from source of exposure. If irritation or signs of toxicity occur, seek medical attention.
- **After skin contact:**
Take off any clothing that the product touched.
Rinse skin with running water for 15 to 20 minutes. Seek medical attention if irritation or signs of toxicity occur.
- **After eye contact:**
Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. If eye irritation persists, get medical advice/attention. Wash hands after handling.
- **After swallowing:** Rinse mouth with water. If irritation or signs of toxicity occur, seek medical attention.
- **Information for Medical Personnel**
 - **Most important symptoms and effects, both acute and delayed:** None expected
 - **Medical conditions aggravated by exposure:** None known

5 Fire-fighting measures

- **Suitable extinguishing agents**
Dry chemical, carbon dioxide (CO₂), water spray or regular foam.
 - Caution: CO₂ will displace air in confined spaces and may cause an oxygen-deficient atmosphere.
 - For larger fires: There are no unique chemical or reactivity hazards that would impact firefighting decisions related to this product. Use firefighting measures that suit the environment.
- **Special hazards arising from the substance or mixture**
There are no unique chemical or reactivity hazards that would impact firefighting decisions due to the chemicals in this product.
- **Protective equipment**
For large fires, wear appropriate heat- and flame-resistant personal protective equipment and a NFPA/NIOSH approved positive-pressure, self-contained breathing apparatus.

6 Accidental release measures

- **Personal precautions, protective equipment and emergency procedures**
Minimize exposure by using appropriate personal protective equipment as listed in Section 8. Stop leak if possible. Keep unprotected persons away.
- **Environmental precautions**
Prevent liquid and vapor from entering sewage system, storm drains, surface waters, and soil.



Abbott

© Abbott Laboratories Release date 05/30/2018

Safety Data Sheet

Last alteration on 05/03/2018

Product name: Alinity c ICT Urine Calibrator Low

· **Methods and material for containment and cleaning up**

Blot up small volumes of spilled or spattered product with paper towels or similar materials.

- Contain larger spills by placing absorbants around the outside edges of the spill. Absorb with any material suitable for water-based liquids - e.g. paper towels, universal sorbents, sand, diatomite, sawdust, etc.

Clean the affected area. Suitable cleaners are:

- warm water and detergent or similar cleansing agent

Dispose of spilled and contaminated material in accordance with Federal, State, and Local regulations. See Section 13 for information that may impact disposal of materials contaminated with this product.

· **Reference to other sections**

See Section 7 for information on safe handling.

See Section 8 for information on personal protection equipment.

See Section 13 for disposal information.

7 Handling and storage

· **Precautions for safe handling:**

Avoid direct contact with material. If handled, wash thoroughly. Practice general safety precautions.

· **Information about protection against explosions and fires:** The product is not flammable.

· **Requirements to be met by storerooms and receptacles:**

Store only in the original container.

Refer to the package insert or product label for additional information on storage conditions for product quality.

· **Information about storage in one common storage facility:** Store in original packaging.

· **Further information about storage conditions:** Protect from heat and direct sunlight.

8 Exposure controls/personal protection

· **Components with Occupational Exposure Limits**

The product does not contain any hazardous ingredients with occupational exposure limits established by OSHA, ACGIH, or NIOSH.

· **General protective and hygienic measures:**

Always maintain good housekeeping and follow general precautionary measures. Do not eat, drink or store food and beverages in areas where chemicals or specimens are used. Wash hands before breaks, after handling reagents and specimens, and at the end of the workshift.

· **Breathing equipment:**



Product name: Alinity c ICT Urine Calibrator Low

Normal use and storage of product - respiratory protection is not necessary if room is well ventilated.

Small-volume spills (e.g. small enough to clean up with a paper towel or small sorbent pad) - respiratory protection should not be necessary if room is well ventilated.

Other unusual conditions (e.g. volume spilled too big to clean up with materials in arm's reach) - Use appropriate NIOSH-approved air-purifying respirator if airborne chemical concentrations may exceed the exposure limit (if any) listed above.

Hazardous Materials Emergencies or Firefighting - use NIOSH/NFPA-approved respiratory protection.

· **Hand protection:**

Wear impervious gloves if hand contact with the material is anticipated. Dispose of contaminated gloves after use in accordance with applicable laws and good laboratory practices.

· **Material of gloves and breakthrough time of the glove material:**

The glove material must be suitable for use in a microbiological laboratory and have a measured breakthrough time of at least 30 minutes, such as those with a Class 2 protection index per EN374 (or equivalent standard applicable in your region). NOTE: This recommendation applies only to the product stated in this Safety Data Sheet. When dissolving in or mixing with other substances, contact the supplier of approved gloves.

· **Eye protection:**

Wear safety glasses or other protective eyewear. If splash potential exists, wear full face shield or goggles.

· **Body protection:**

Normal use: protect personal clothing from splatters and small spills. Wear a laboratory coat (or other protective clothing required by your institution).

Larger spills (e.g. that can saturate cloth): wear appropriate water-repellant covering over clothing.

9 Physical and chemical properties

· **General Information**

· **Form:** Solution
· **Color:** Light yellow

· **Odor:** Odorless

· **pH-value at 20 °C (68 °F)** 7
· **Melting point/Melting range:** Not determined
· **Boiling point/Boiling range:** Not determined

· **Flash point** Not applicable
· **Flammability (solid, gaseous)** Not applicable
· **Auto igniting** Product is not self-igniting.

· **Danger of explosion** Product does not present an explosion hazard.

· **Explosion limits**
· **Lower:** Not determined
· **Upper:** Not determined

· **Density at 20 °C (68 °F)** 1 g/cm³ (8.345 lbs/gal)



Product name: Alinity c ICT Urine Calibrator Low

· Evaporation rate:	Not determined
· Solubility in / Miscibility with	
· Water:	Fully miscible
· Dynamic:	Not determined
· Water:	99.6 %
· Solids content:	0.0 %

10 Stability and reactivity

- **Thermal decomposition / conditions to be avoided**
No decomposition if used and stored according to specifications.
- **Possibility of hazardous reactions:** No dangerous reactions known.
- **Conditions to avoid:** No further relevant information available.
- **Incompatible materials:** No further relevant information available.
- **Hazardous decomposition products:** No dangerous decomposition products known.

11 Toxicological information

- **Acute toxicity**
 - **LD50/LC50 values for hazardous ingredients per OSHA criteria:**
 - **Ingredients (100% pure substance/s):** Not applicable.
 - **Primary toxicological effects of the final product:**
 - **Skin irritation:** No irritant effect.
 - **Eye irritation:** No irritant effect.
 - **Sensitization:** No sensitizing effects known.
- **Additional toxicological information:** None
 - **Carcinogenic categories**

· IARC (International Agency for Research on Cancer)
None of the ingredients is listed.

· NTP (National Toxicology Program)
None of the ingredients is listed.

· OSHA-Ca (Occupational Safety & Health Administration)
None of the ingredients is listed.

- **Target organs/systems:** Unknown



Product name: Alinity c ICT Urine Calibrator Low

12 Ecological information

- **Aquatic toxicity:** No further relevant information available.
- **Additional ecological information**
 - **General notes:** Not known to be a water pollutant.
- **Results of PBT and vPvB assessment**
 - **PBT:** Not applicable
 - **vPvB:** Not applicable

13 Disposal considerations

- **Recommendation for disposal of unused product:**
Dispose in accordance with federal, state and local regulations.
- **Recommendation for disposal of packaging:**
Non-contaminated packaging may be used for recycling. Refer to applicable local regulations and institutional policies.
For disposal of contaminated packaging, refer to applicable local regulations and institutional policies.
- **Recommended cleansing agent:** Water with cleansing agents, if necessary.

14 Transport information

· DOT, ADN, IMDG, IATA none

· UN proper shipping name

· DOT, ADR, ADN, IMDG, IATA none

· Transport hazard class(es)

· DOT, ADR, ADN, IMDG, IATA

· Class none

· DOT, IMDG, IATA none

· Environmental hazards

· Marine pollutant: No

· Additional information

· DOT

· Remarks: Not restricted for transportation.

· ADR

· Remarks: Not restricted for transportation.

· IMDG

· Remarks: Not restricted for transportation.



Abbott

© Abbott Laboratories Release date 05/30/2018

Safety Data Sheet

Last alteration on 05/03/2018

Product name: Alinity c ICT Urine Calibrator Low

· IATA

· **Remarks:** Not restricted for transportation.

15 Regulatory information

· **SARA (Superfund Amendments and Reauthorization Act of 1986 - USA):**

· **Section 302/304 (40CFR355.30 / 40CFR355.40):**

The product does not contain listed substances.

· **Section 313 (40CFR372.65):**

The product does not contain listed substances.

· **California Proposition 65 (USA):**

· **Chemicals known to cause cancer:**

The product does not contain listed substances.

· **Chemicals known to cause female reproductive toxicity:**

None of the ingredients is listed.

· **Chemicals known to cause male reproductive toxicity:**

None of the ingredients is listed.

· **Chemicals known to cause developmental reproductive toxicity:**

None of the ingredients is listed.

16 Other information

The information and recommendations contained herein are based upon information or tests believed to be reliable. Abbott Laboratories does not guarantee the accuracy or completeness of this information or recommendations contained herein, NOR SHALL ANY OF THIS INFORMATION CONSTITUTE A WARRANTY, WHETHER EXPRESSED OR IMPLIED, AS TO THE SAFETY OF THE GOODS, THE MERCHANTABILITY OF THE GOODS, OR THE FITNESS OF THE GOODS FOR A PARTICULAR PURPOSE.

This information is not a substitute for the advice of a health care professional, nor is it a recommendation for any particular course of treatment. It is not intended to supplement, modify or supersede any information (e.g. labeling and package inserts) provided with respect to the medical use of the product. Abbott Laboratories assumes no responsibility for results obtained or for incidental or consequential damages, including lost profits, arising from the use of these data. No warranty against infringement of any patent, copyright or trademark is made or implied.

· **Department issuing SDS**

- Abbott Diagnostics Safety, Health and Environmental Assurance
Department 0571

· **Contact**

- General information about this product:
Abbott Diagnostics
Technical Support
100 Abbott Park Road
Abbott Park, IL 60064-3500



Abbott

© Abbott Laboratories Release date 05/30/2018

Safety Data Sheet

Last alteration on 05/03/2018

Product name: Alinity c ICT Urine Calibrator Low

Phone: 1-877-4 ABBOTT

· **Date of preparation / last revision** 05/30/2018 / 3

· **Abbreviations and acronyms:**

RID: Règlement international concernant le transport des marchandises dangereuses par chemin de fer (Regulations Concerning the International Transport of Dangerous Goods by Rail)

ICAO: International Civil Aviation Organisation

ADR: Accord européen sur le transport des marchandises dangereuses par Route (European Agreement concerning the International Carriage of Dangerous Goods by Road)

IMDG: International Maritime Code for Dangerous Goods

DOT: US Department of Transportation

IATA: International Air Transport Association

EINECS: European Inventory of Existing Commercial Chemical Substances

ELINCS: European List of Notified Chemical Substances

CAS: Chemical Abstracts Service (Division of the American Chemical Society)

LC50: Lethal concentration, 50 percent

LD50: Lethal dose, 50 percent

PBT: persistent, bioaccumulative and toxic

vPvB: very persistent and very bioaccumulative

OSHA: Occupational Safety & Health

TLV: Threshold Limit Value

PEL: Permissible Exposure Limit

REL: Recommended Exposure Limit

· * **Sections marked with an asterisk (*) have been altered since the previous version.**