



Medline Industries, Inc.
Three Lakes Dr.
Northfield, IL 60093
1.630.865.6288
1.800.MEDLINE (633.5463)

CERTIFICATE OF ANALYSIS

**MCHEM140 - Isopropyl Alcohol 99%, 1
gal (MTO)**

Lot Number: 248062

Expiration Date: 2027-December-31

COMPONENTS

Test	Specifications	Results
Color	Clear	Clear
Water (%)	≥ 0.00 and ≤ 0.10	0.04
APHA (Pt-Co units)	≥ 0.00 and ≤ 5.00	3.00
Isopropanol (%)	≥ 99.50 and ≤ 100.00	100.00
Refractive Index @ 20C (nD)	≥ 1.36 and ≤ 1.38	1.38
Specific Gravity @ 20C (g/cm ³)	≥ 0.785 and ≤ 0.788	0.785

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control - This report is electronically produced and is valid without signature.

Source: MCHEM140/248062
Form: WM2103_V5