



Medline Industries, Inc.
Three Lakes Dr.
Northfield, IL 60093
1.630.865.6288
1.800.MEDLINE (633.5463)

CERTIFICATE OF ANALYSIS

MCHEM140 - Isopropyl Alcohol 99%, 1 gal (MTO)

Lot Number: 195627

Expiration Date: 2026-April-30

COMPONENTS

Test	Specifications	Results
Color	Clear	Clear
Water (%)	≥ 0.00 and ≤ 0.10	0.03
APHA (Pt-Co units)	≥ 0.00 and ≤ 5.00	4.00
Isopropanol (%)	≥ 99.50 and ≤ 100.00	100.00
Refractive Index @ 20C (nD)	≥ 1.36 and ≤ 1.38	1.38
Specific Gravity @ 20C (g/cm ³)	≥ 0.79 and ≤ 0.79	0.79

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control - This report is electronically produced and is valid without signature.

Source: MCHEM140/195627
Form: WM2103_V5